



UNIVERSITÀ
degli STUDI
di CATANIA

Test Mascherine Chirurgiche

G.Cuttone, L. D'Antoni, A. Miraglia, S. Musumeci, S. Passarello

INFN-LNS

S. Baglio, S. Stefani, S. Stracquadanio, N. Tuccitto

Università di Catania

UNI EN 14683

NORMA
EUROPEA

**Maschere facciali ad uso medico - Requisiti e metodi di
prova**

UNI EN 14683

OTTOBRE 2019

Medical face masks - Requirements and test methods

La norma specifica i requisiti di fabbricazione, progettazione e prestazione e i metodi di prova per le maschere facciali ad uso medico destinate a limitare la trasmissione di agenti infettivi tra pazienti e personale clinico durante gli interventi chirurgici e altri contesti medici con requisiti simili.

Table 1 — Performance requirements for medical face masks

Test	Type I ^a	Type II	Type IIR
Bacterial filtration efficiency (BFE), (%)	≥ 95	≥ 98	≥ 98
Differential pressure (Pa/cm ²)	< 40	< 40	< 60
Splash resistance pressure (kPa)	Not required	Not required	≥ 16,0
Microbial cleanliness (cfu/g)	≤ 30	≤ 30	≤ 30

^a Type I medical face masks should only be used for patients and other persons to reduce the risk of spread of infections particularly in epidemic or pandemic situations. Type I masks are not intended for use by healthcare professionals in an operating room or in other medical settings with similar requirements.

B.7.1 Assemble the test apparatus in accordance with the flow chart shown in Figure B.1 or Figure B.3.





Figure B.1 — Principle of BFE test apparatus

B.7.3. **AC** The test area shall be minimum 49 cm² **AC**. Alternative means to position the sample may be appropriate, but, if deviated from the procedure, this shall be documented in the test report.



Key

- 1 inlet cone
- 2 o'ring inlet cone
- 3 cloth / mask

Figure B.2 — Placement of test specimen on the cascade impactor

Stage arrangement

Stage number	1	2	3	4	5	6
Size of particle	P1	P2	P3	P4	P5	P6
Viable "particle" plate count	C1	C2	C3	C4	C5	C6

where

$$P1 = 7,00 \mu\text{m}$$

$$P2 = 4,70 \mu\text{m}$$

$$P3 = 3,30 \mu\text{m}$$


$$P4 = 2,10 \mu\text{m}$$

$$P5 = 1,10 \mu\text{m}$$

$$P6 = 0,65 \mu\text{m}$$



$$MPS = \frac{(P1 \times C1) + (P2 \times C2) + (P3 \times C3) + (P4 \times C4) + (P5 \times C5) + (P6 \times C6)}{C1 + C2 + C3 + C4 + C5 + C6}$$




(B.1)








Key



- | | | | |
|---|--------------------------|----|--------------------------|
| 1 | drive mechanism | 8 | cascade impactor |
| 2 | bacterial suspension | 9 | outlet to sink |
| 3 | nebulizer | 10 | condenser |
| 4 | filter | 11 | cold water inlet |
| 5 | aerosol chamber | 12 | calibrated flow meter |
| 6 | high pressure air source | 13 | compressor (vacuum pump) |
| 7 | test material | | |




















UNIVERSITÀ
degli STUDI
di CATANIA

Test Respirabilità

UNI EN 14683

A device which measures the differential pressure required to draw air through a measured surface area at a constant air flow rate is used to measure the air exchange pressure of the medical face mask material, as shown in Figure C.1. A water-filled (or digital) differential manometer is used to measure the differential pressure. A mass flow meter is used for measurement of the airflow. An electric vacuum pump draws air through the test apparatus and a needle valve is used to adjust the airflow rate.



Key

- | | | | |
|---|--|----|--|
| 1 | air inlet | 7 | metallic ring (3 mm thick) |
| 2 | mass flow meter | 8 | filter material |
| 3 | lever for mechanical clamping | 9 | differential manometer or M1 and M2 manometers |
| 4 | system for final adjustment of the pressure
(either at the top or the bottom) | 10 | valve |
| 5 | system ensuring optimal alignment of the 2 parts
of the sample holder | 11 | vacuum pump including a pressure buffer tank |
| 6 | sample holder with a metal sealing mechanism | 12 | mass flow meter for checking leaks (optional) |

Figure C.1 — Test apparatus for measuring differential pressure





Alcuni Numeri

- 254 aziende
- Oltre 400 tipologie di mascherine testate
- Solo il 5% hanno superato i test

Count of REGIONE

Abruzzo

3,5%

Veneto

2,3%

Molise

1,9%

Lombardia

3,5%

Toscana

3,1%

Piemonte

2,3%

Puglia

7,8%

Calabria

6,6%

Campania

10,9%

254 aziende

Oltre 400 tipologie di mascherine testate

Solo il 5% hanno superato i test

